



Product Service

CERTIFICATE

No. Z1 06 05 42940 016

Holder of Certificate: Invacare EC-Hong A/S

Ostergade 3
4270 Hong
DENMARK

Certification Mark:



Product:

Care beds

The product was tested on a voluntary basis and complies with the essential requirements. The certification mark shown above can be affixed on the product. See also notes overleaf.

Test report no.:

71306788



Date, 2006-06-08

Page 1 of 2



Product Service

CERTIFICATE

No. Z1 06 05 42940 016

Model(s):**Electrically adjustable care bed
Type: Etude Medley****Parameters:**

total length:	2215 mm
total width:	1020 mm
mattress support platform:	2000 mm x 900 mm, four sections
controller:	CB 6444 + 10019, co. Linak
motor stroke rack:	2 x LA 27, 1500 N, co. Linak
motor backrest:	LA 27, 2500 N, co. Linak
motor leg section:	LA 27, 6000 N, co. Linak
hand control:	HB7X001-01, IP66, co. Linak
locking box:	ACL 001-0114238644, IP 54, co. Linak
protection class:	II
degree of protection:	IP 54
transformer:	Uin: 230V~, 50 Hz, lin: max. 1 A, max. 70 VA
cyclic duration factor:	10%, max. 2/18 min.
height adjustable:	approx. 400 to 810 mm
max. patient weight:	145 kg
max. load:	180 kg
side rails:	metal, Scala basic (1432784-0152); metal, Medley version 4 parts (1491226-0152); wood, Medley version 4 parts (1491224-0101)
mattress:	90 x 200 x 10/12 cm

Tested according to:

EN 1970:2000
EN 60601-2-38/A1:2000
(with regard to / in Anlehnung)
Aus der EN 60601-2-38/A1:2000 wurden nur die für die häusliche Pflege relevanten Prüfungen herangezogen.
Only the relevant parts concerning the home-care were quoted from the EN 60601-2-38/A1:2000.
Die Trendelenburglagerung und die Anti-Trendelenburglagerung sind nicht einstellbar.
Trendelenburgpositioning and Anti-Trendelenburgpositioning are not adjustable.

Production Facility(ies):

42940

Page 2 of 2